

ONCORAL

Instruction for Use

Kit Contents :

1 extraction solution tube (Blue Cap), 1 Beaker (Orange Lid) and 2 sample tubes (Orange caps)

Hardware requirement not provided :

Clock or Stop watch to measure time

Storage and shelf life :

Storage: safe from the light, at a temperature ranging between 4°C and 25°C.

Shelf life: See the labels on tubes.

The extraction solution must be used immediately after opening. Once opened, the tube cannot be used later.

General recommendations concerning the saliva sampling :

No food or drink must be consumed 10 minutes before the saliva sampling.

Procedure :



Open the **extraction solution tube** and **Rinse** the oral cavity with its contents **during 2 minutes**.



Open the **Beaker** and **spit** there all the contents located in mouth. **Close the beaker**.



Withdraw the round adhesive of safety on the beaker. **Place the beaker on a plane surface** Take the **Sample tube (NOT TO OPEN It)** and **firmly insert** it in the opening with the cap to the bottom. The **sample tube** fills automatically. Bring out the **sample tube** as soon as it is filled. Reiterate the same operation with the second sample tube. If it remains liquid in the **beaker**, the sample and the beaker must be eliminated Homogenize the contents of the **sample tubes** and preserve at **room temperature** until forwarding. After use, the cover in the lid of the **beaker** must be closed again using the round adhesive of safety.



Warning statements and precautions :

A not blocked nasal breathing represents the prerequisite for saliva sampling with the system. **Extraction solution tube** : The saliva extraction solution contains food dye FD&C yellow n° 5 (tartrazine). In the event of accidental absorption of the solution, no health risk, and it is not necessary to inform the doctor. In rare cases, the FD&C yellow n° 5 (tartrazine) can start allergic reactions, in particular for people that have a sensitivity to aspirine and/or benzoate like at the asthmatic ones. In the event of a known intolerance, the use is disadvised.

Beaker : The round adhesive of safety placed on the lid should be withdrawn only for the transfer of the solution in the **sample tube**. Then, it must be positioned back. **Never** put the fingers in the opening because of the risk of puncture per needle!

Sample tubes : The tubes contain sodium toxic* azide. These tubes should not be open. In the event of accidental absorption, immediately consult a doctor by showing him the directions for use.

* Specific precautions for the human being and the environment : R 28 Very toxic in the event of absorption. R 32 in contact with an acid, emits a very toxic gas. R 50/53 Very toxic for the watery organizations, can involve harmful effects in the long run for the watery environment.

ONCORAL

Instruction for Use

Kit Contents :

1 extraction solution tube (Blue Cap), 1 Beaker (Orange Lid) and 2 sample tubes (Orange caps)

Hardware requirement not provided :

Clock or Stop watch to measure time

Storage and shelf life :

Storage: safe from the light, at a temperature ranging between 4°C and 25°C.

Shelf life: See the labels on tubes.

The extraction solution must be used immediately after opening. Once opened, the tube cannot be used later.

General recommendations concerning the saliva sampling :

No food or drink must be consumed 10 minutes before the saliva sampling.

Procedure :



Open the **extraction solution tube** and **Rinse** the oral cavity with its contents **during 2 minutes**.



Open the **Beaker** and **spit** there all the contents located in mouth. **Close the beaker**.



Withdraw the round adhesive of safety on the beaker. **Place the beaker on a plane surface** Take the **Sample tube (NOT TO OPEN It)** and **firmly insert** it in the opening with the cap to the bottom. The **sample tube** fills automatically. Bring out the **sample tube** as soon as it is filled. Reiterate the same operation with the second sample tube. If it remains liquid in the **beaker**, the sample and the beaker must be eliminated Homogenize the contents of the **sample tubes** and preserve at **room temperature** until forwarding. After use, the cover in the lid of the **beaker** must be closed again using the round adhesive of safety.



Warning statements and precautions :

A not blocked nasal breathing represents the prerequisite for saliva sampling with the system. **Extraction solution tube** : The saliva extraction solution contains food dye FD&C yellow n° 5 (tartrazine). In the event of accidental absorption of the solution, no health risk, and it is not necessary to inform the doctor. In rare cases, the FD&C yellow n° 5 (tartrazine) can start allergic reactions, in particular for people that have a sensitivity to aspirine and/or benzoate like at the asthmatic ones. In the event of a known intolerance, the use is disadvised.

Beaker : The round adhesive of safety placed on the lid should be withdrawn only for the transfer of the solution in the **sample tube**. Then, it must be positioned back. **Never** put the fingers in the opening because of the risk of puncture per needle!

Sample tubes : The tubes contain sodium toxic* azide. These tubes should not be open. In the event of accidental absorption, immediately consult a doctor by showing him the directions for use.

* Specific precautions for the human being and the environment : R 28 Very toxic in the event of absorption. R 32 in contact with an acid, emits a very toxic gas. R 50/53 Very toxic for the watery organizations, can involve harmful effects in the long run for the watery environment.